



PUDUCHERRY POWER CORPORATION LIMITED
(A GOVERNMENT OF PUDUCHERRY UNDERTAKING)
T.R PATTINAM -609606, KARAİKAL



Phone No: 04368 - 233060, 233988

mail : ppcl.kkl@nic.in.

Fax: 04368- 233355

Ref. No: CP-274/PUR/2024-25/1611

GSTIN No.: 34AAACP6507E1ZS

Dt. 03.12.2024.

ENQUIRY

To

Sub: Inviting offer pH & Conductivity meter - Reg.

QUOTATION TO BE

Submitted on or before 26.12.2024 - 4.00 PM	Opened on 27.12.2024 - 11.00 AM
Kept valid for 3 months from the date of opening	Submitted in one part.

Sealed quotations are invited for the following item as per our terms and conditions mentioned therein.

Description	Specification	Qty.
1. pH meter	As per annexure (Make : Deep Vision/ Electronics India/Systronics/Labtronics)	1 No.
2. Conductivity meter		1 No.

Terms and conditions:

1. The rate quoted should be firm and inclusive of all taxes and other charges as admissible.
2. The materials should be directly supplied to the PPCL site, T.R.Pattinam, Karaikal - 609 606, on door delivery basis.
3. Payment will be made within 30 days from the date of receipt of your Invoice after receipt of materials at site in good condition. **The bill in triplicate along with advance stamped receipt should be sent to the undersigned for effecting payment.**
4. No payment will be made against document through bank. Part payment will not be accepted for the supply of ordered items.
5. Applicable taxes and charges to be stated individually - HSN no. to be stated
6. Our GSTIN No.: 34AAACP6507E1ZS and PAN No.: AAACP6507E
7. Any loss or damage during transit to site should be made good by the suppliers free of charge provided the same is reported within 15 days of the arrival of material at site.
8. The delivery period and mode of despatch should be clearly specified.
9. If the material is rejected for quality reasons, then the supplier shall at his own cost, take back and replace the same.
10. One year warranty from the date of supply, for the meters shall be provided and should be serviced/replaced at sellers cost.
11. Sealed quotations should be super scribed as "QUOTATION/ENQUIRY NO.CP-274/PUR/2024-25 for the supply of pH & Conductivity meter due on 26/12/2024".
12. If the tender opening date happens to be a holiday, the tender will be opened on the next working day.
13. The undersigned reserves all rights to accept/reject any or all the quotations without assigning any reasons thereof.

03/12/2024

7/2

EXECUTIVE ENGINEER(Mech.) Plant



PUDUCHERRY POWER CORPORATION LIMITED

(A GOVERNMENT OF PUDUCHERRY UNDERTAKING)

T.R PATTINAM -609606, KARAIKAL



Phone No: 04368 - 233060, 233988

mail : ppcl.kkl@nic.in

Fax: 04368- 233355

Ref. No: CP-274/PUR/2024-25/

GSTIN No.: 34AAACP6507E1ZS

Dt. 03.12.2024.

ENQUIRY PAGE NO. 2

(This page shall be treated as bid, if related entries are made and duly signed by the party and returned to PPCL)

Description	Qty Reqd.	Total Amount (Rs.)
pH Meter, Electronic display type, Table top, range 0-14 pH, resolution : 0.01, Auto temp. compensation, 3 1/2 digit display, 230 V power supply, One No. Electrode stand with clamp, pH Electrode (dip type) (Make : Deep Vision/ Electronics India/Systronics/Labtronics)	1 No.	
Conductivity Meter, Electronic display type, Table top, range 200 µ Mhos/cm to 1000 mMhos/cm, resolution : 1.1 µMhos/cm, Auto temp. compensation, Cell constant : 0.4 to 1.5 adjustable, 3 1/2 digit display, 230 V power supply, One No. Electrode stand with clamp, Conductivity Electrode (dip type) (Make : Deep Vision/ Electronics India/Systronics/Labtronics)	1 No	
Add: P&F Charges (if any)		
Add: Freight Charges to PPCL site (if any)		
Add: GST@ %		
Total		

Commercial Terms and Conditions:

Sl.No	Details	Terms in Tender	Terms Accepted by the Party
1	Price Quoted	F.O.R. PPCL Site, T.R.Pattinam, Karaikal	
2	Payment Terms	Within 30 days after delivery of the materials.	
3	Validity	90 Days from date of opening of Tender	
4	Warranty	Twelve months from the date of supply	
4	Delivery Period	Vendor to specify	
5	Acceptance of All the Terms and Condition of the Enquiry	To be Accepted	

Signature with Seal of the Party



PUDUCHERRY POWER CORPORATION LIMITED

(A GOVERNMENT OF PUDUCHERRY UNDERTAKING)

T.R PATTINAM -609606, KARAIKAL



Phone No: 04368 - 233060, 233988

mail : ppcl.kkl@nic.in.

Fax: 04368- 233355

GSTIN No.: 34AAACP6507E1ZS

Dt. 03.12.2024.

Ref. No: CP-274/PUR/2024-25/

Annexure to Enquiry : Ref. No: CP-274/PUR/2024-25/

Dt. 03.12.2024.

Sl.No.	Description	pH meter	Conductivity meter
1	Range	0 -14 pH mV range : 0 to \pm 1999 mV	200 μ Mhos/cm to 1000 mMhos/cm
2	Accuracy	pH \pm 0.01 pH \pm 1 digit mV: \pm 1 mV \pm 1 digit	\pm 1 % F.S. \pm 1 digit
3	Resolution	0.01pH : 1 mv	1.1 μ Mhos/cm
4	Temp. compensation	0 to 100 Deg C (Auto mode)	0 to 50 Deg C (Auto mode)
5	Input Impedance	$> 10^{13}$ ohms	
5	Cell Constant	-	0.4 to 1.5 adjustable
6	Display	3 ¹ / ₂ Digit bright red LED display	
7	Power supply	230 V \pm 10% AC, 50 Hz, BNC connection	
8	Accessories	One No. Electrode stand with clamp, pH Electrode (dip type)	One No. Electrode stand with clamp, Conductivity cell (dip type)
9	Make	(Make : Deep Vision/ Electronics India/Systronics /Labtronics)	